

VISCOSITY CUPS ACCORDING TO BS3900 (1971) OLD SPECIFICATION

SH0070, SH0127, SH0165, SH0201, SH0230

PRODUCT DESCRIPTION

The BS Viscosity Cups is a range of anodised aluminium viscosity cups with integral nozzle. Suited for measuring paints, lacquers and other fluids. (Note: the cup doesn't have a stainless steel insert).

BUSINESS

Laboratory

STANDARDS

According to (old type) BS3900: Part A6, 1971. This specification is replaced by BS EN ISO 2431/853900 part A6:1996 –(see BS EN ISO 2431 series of cups).

ORDERING INFORMATION

SH0070	BS Flow Cup, BS3900 (1971) old specification (B2)
SH0127	BS Flow Cup, BS3900 (1971) old specification (B3)
SH0165	BS Flow Cup, BS3900 (1971) old specification (B4)
SH0201	BS Flow Cup, BS3900 (1971) old specification (B5)
SH0230	BS Flow Cup, BS3900 (1971) old specification (B6)

ACCESSORIES

VF2061	Tripod stand Type S40B, stainless steel ring incl. Spirit level
DI0076	Stopwatch Type C510 digital LCD-display, 9h. 59 min. 59,99 sec.

SPECIFICATIONS

Code	Orifice diameter	Viscosity range	Flow times
SH0070 - B2	2.38mm (0.09")	38-71cSt	30-300 secs
SH0127 - B3	3.17mm (0.12")	38-147cSt	
SH0165 - B4	3.97mm (0.16")	71-455cSt	
SH0201 - B5	4.76mm (0.19")	299-781cSt	
SH0230 - B6	7.14mm (0.28")	781-1650cSt	

SPECIAL CARE

- A viscosity cup is a precision instrument. With reasonable care, it is constructed to give many years of satisfactory service. To clean the instrument, use a soft cloth, NEVER clean by any mechanical means, such as sandpaper, steel brush or any other abrasive tool.



- Particular care should be used in cleaning the orifice to avoid leaving deposits or scratches on internal surfaces.
- It's recommended to clean the cup promptly after each use, unless it will be used immediately for a rerun of the same material.

SAFETY PRECAUTIONS

Determining viscosity may involve hazardous materials, operations and equipment. It is the responsibility of the executor to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to the measurement.

DISCLAIMER

The right of technical modifications is reserved.

The information given in this sheet is not intended to be exhaustive and any person using the product for any purpose other than that specifically recommended in this sheet without first obtaining written confirmation from us as to the suitability of the product for the intended purpose does so at his own risk. Whilst we endeavour to ensure that all advice we give about the product (whether in this sheet or otherwise) is correct we have no control over either the quality or condition of the product or the many factors affecting the use and application of the product. Therefore, unless we specifically agree in writing to do so, we do not accept any liability whatsoever or howsoever arising for the performance of the product or for any loss or damage (other than death or personal injury resulting from our negligence) arising out of the use of the product. The information contained in this sheet is liable to modification from time to time in the light of experience and our policy of continuous product development.